

To protect the privacy of customers or suppliers of Salamander Jewelry.
We erased their address from this document.

M/C : SHINVA

MODEL : XG 1 SERIES

SERIAL No. : 070997

Document No. : TR-EO-033060516

Date : 6/5/2016

Page : 2 / 4

PRODUCT : ORDER NO. 2145876

SAMPLE SLIP PARAMETER

EXHAUST:04H DOWN TOTAL :11h.56min
PRETT :00H 50MIN STERI :07H 00MIN
ACTUAL TIME

18:05:38 STEILIZATION SUCCESS

END	18:05:27	058,80	-005kPa
	18:05:27	058,80	-005kPa
	14:08:39	059,60	-015kPa
EXHAUST	13:58:27	059,60	-001kPa
	13:56:57	059,40	-001kPa
	13:46:57	059,50	-002kPa
	13:36:57	059,60	-001kPa
	13:26:57	059,30	-001kPa
	13:16:57	059,50	-001kPa
	13:06:57	059,50	-001kPa
	12:56:57	059,40	-002kPa
	12:46:57	059,50	-001kPa
	12:36:57	059,30	000kPa
	12:26:57	059,30	-001kPa
	12:16:57	059,40	-001kPa
	12:06:57	059,40	-001kPa
	11:56:57	059,40	-001kPa
	11:46:57	059,20	-001kPa
	11:36:57	059,10	-001kPa
	11:26:57	058,90	-001kPa
	11:16:57	059,00	-001kPa
	11:06:57	059,00	-001kPa
	10:56:57	058,90	-001kPa
	10:46:57	059,00	-001kPa
	10:36:57	058,80	-001kPa
	10:26:57	058,70	-001kPa
	10:16:57	058,60	-001kPa
	10:06:57	058,30	-001kPa

	09:56:57	058,20	-001kPa
	09:46:57	058,40	-001kPa
	09:36:57	058,10	-002kPa
	09:26:57	058,10	-001kPa
	09:16:57	057,90	-002kPa
	09:06:57	057,90	-002kPa
	08:56:57	057,70	-001kPa
	08:46:57	057,60	-001kPa
	08:36:57	057,50	-001kPa
	08:26:57	057,10	-002kPa
	08:16:57	057,10	001kPa
	08:06:57	056,80	-001kPa
	07:56:57	056,50	000kPa
	07:46:57	056,10	000kPa
	07:36:57	055,70	000kPa
	07:26:57	055,90	000kPa
	07:16:57	055,90	001kPa
	07:06:57	056,10	117kPa
STERI	06:59:13	056,40	311kPa
	06:56:57	056,20	003kPa
ADD EO	06:53:28	056,20	-076kPa
	06:46:57	056,20	-015kPa
HUMIDI	06:43:26	055,90	-074kPa
PRETT	06:36:57	055,90	-001kPa
PRHEAT	06:09:10	055,40	000kPa
STATE	TIME	TEMP	PRE

RUN COUNT : 0666

EO PRE :060kPa VENT TIME:04h 00m

STERI TEMP: 056,00 STERI TIME:07h 00m

PRO TYP 06:09:10 055,40 000kPa

START TIME: 2016-05-04 06:09:02