

To protect the privacy of customers or suppliers of Salamander Jewelry.  
We erased their address from this document.

SAMPLE SLIP PARAMETER

EXHAUST:04H 00MIN TOTAL :11H 54MIN  
 PRETT :00H 51MIN STERI :07H 00MIN  
 ACTUAL TIME

---

21:46:17 STEILIZATION SUCCESS

END	21:46:06	059,50	-005kPa
	21:46:06	059,50	-005kPa
	17:50:18	061,60	-014kPa
EXHAUST	17:43:06	061,60	000kPa
	17:41:36	061,70	000kPa
	17:31:36	061,70	000kPa
	17:21:36	061,80	000kPa
	17:11:36	061,90	000kPa
	17:01:36	061,80	000kPa
	16:51:36	062,00	000kPa
	16:41:36	062,00	000kPa
	16:31:36	062,30	001kPa
	16:21:36	062,50	000kPa
	16:11:36	062,00	000kPa
	16:01:36	062,10	000kPa
	15:51:36	061,80	000kPa
	15:41:36	061,80	000kPa
	15:31:36	062,20	000kPa
	15:21:36	062,20	000kPa
	15:11:36	061,90	-001kPa
	15:01:36	062,00	-001kPa
	14:51:36	061,90	001kPa
	14:41:36	061,80	001kPa
	14:31:36	061,70	000kPa
	14:21:36	061,70	001kPa
	14:11:36	061,70	000kPa
	14:01:36	061,60	000kPa
	13:51:36	061,70	000kPa

	13:41:36	061,50	002kPa
	13:31:36	061,50	000kPa
	13:21:36	061,30	000kPa
	13:11:36	061,30	000kPa
	13:01:36	061,40	001kPa
	12:51:36	061,10	-001kPa
	12:41:36	061,20	000kPa
	12:31:36	061,20	000kPa
	12:21:36	061,10	000kPa
	12:11:36	060,90	000kPa
	12:01:36	060,90	000kPa
	11:51:36	060,80	055kPa
	11:41:36	060,80	064kPa
	11:31:36	060,70	076kPa
	11:21:36	060,60	089kPa
	11:11:36	060,50	145kPa
	11:01:36	060,60	265kPa
	10:51:36	060,50	311kPa
STERI	10:43:32	060,60	224kPa
	10:41:36	060,70	003kPa
ADD EO	10:37:46	060,70	-080kPa
	10:31:36	060,60	-013kPa
HUMIDI	10:27:44	060,60	-079kPa
PRETT	10:21:36	060,60	001kPa
PRHEAT	09:52:23	060,50	-002kPa
STATE	TIME	TEMP	PRE

---

RUN COUNT : 0768  
 EO PRE :060kPa VENT TIME:04H 00M  
 STERI TEMP: 056,00 STERI TIME:07H 00M  
 PRO TIP 09:52:23 060,50 -002kPa  
 START TIME: 2016-07-17 09:52:15  
 08:32:06 REPOWER