

To protect the privacy of customers or suppliers of Salamander Jewelry.
We erased their address from this document.

M/C : SHINVA

Document No. : TR-EO-004110816

MODEL : XG 1 SERIES

Date : 11/8/2016

SERIAL No. : 070997

Page : 1 / 4

PRODUCT : ORDER NO. 2156187

Determination of Sterilization Specification

Item Detail	Specification	Setting	Result
Chamber Setting Temperature	55.0 - 59.0 (°C)	59 (°C)	-
Vacuum Level	5-12 kPa	9 kPa	-
Vacuum Time (min)	50-60 min	60 min	-
Humidification			
Humidity (%)	50%-90%	87%	-
Sterilization			
Steri Temp	55.0 - 59.0 (°C)	55.0 (°C)	-
Jacket Temp	60.0 - 65.0 (°C)	65.0 (°C)	-
Gas Injection			
Final pressure (kpa(abs))	0.21-0.23 kPa	0.21 kPa	-
Final temperature of vaporizer,Temp 10 (°C)	55.0 - 59.0 (°C)	55 (°C)	-
Exposure			
Temperature at strat point of Exposure time period Temp. 1 (°C)	55.0 - 59.0 (°C)	60 (°C)	-
Temperature at finish point of Exposure time period Temp. 1 (°C)	55.0 - 59.0 (°C)	55 (°C)	-
Exposure Time (min)	390-420 min	420 min	-
Exhaust			
Exhaust Time (hour)	≥ 4 hour	4 hour	-
First Vacuum Level (kPa)	5-12 kPa	9 kPa	-
Second Vacuum Level (kPa)	5-12 kPa	9 kPa	-
Third Vacuum Level (kPa)	5-12 kPa	9 kPa	-
Storage			
Storage Time	≥ 48 hour	≥ 48 hour	-
Chemical tested, Sterile Strip			
EO sterilization temperatures 55-60 C° Time sterile by EO 7 hours .	Brown	Brown	Pass
Biological Test			
EZ Test Biological Indicator	Brown	Brown	Pass
24 hour results for EO	Red	Red	Pass

To protect the privacy of customers or suppliers of Salamander Jewelry.
We erased their address from this document.