

To protect the privacy of customers or suppliers of Salamander Jewelry.  
We erased their address from this document.

Document No. : TR-EO-005190916

Date : 19/9/2016

Page : 1 / 5

PRODUCT : ORDER NO. 2159553

### Determination of Sterilization Specification

Item Detail	Specification	Setting	Result
Chamber Setting Temperature	55.0 - 59.0 (°C)	59 (°C)	-
Vacuum Level	5-12 kPa	9 kPa	-
Vacuum Time (min)	50-60 min	60 min	-
<b>Humidification</b>			
Humidity (%)	50%-90%	87%	-
<b>Sterilization</b>			
Steri Temp	55.0 - 59.0 (°C)	55.0 (°C)	-
Jacket Temp	60.0 - 65.0 (°C)	65.0 (°C)	-
<b>Gas Injection</b>			
Final pressure (kpa(abs))	0.21-0.23 kPa	0.21 kPa	-
Final temperature of vaporizer,Temp 10 (°C)	55.0 - 59.0 (°C)	55 (°C)	-
<b>Exposure</b>			
Temperature at strat point of Exposure time period Temp. 1 (°C)	55.0 - 59.0 (°C)	60 (°C)	-
Temperature at finish point of Exposure time period Temp. 1 (°C)	55.0 - 59.0 (°C)	55 (°C)	-
Exposure Time (min)	390-420 min	420 min	-
<b>Exhaust</b>			
Exhaust Time (hour)	≥ 4 hour	4 hour	-
First Vacuum Level (kPa)	5-12 kPa	9 kPa	-
Second Vacuum Level (kPa)	5-12 kPa	9 kPa	-
Third Vacuum Level (kPa)	5-12 kPa	9 kPa	-
<b>Storage</b>			
Storage Time	≥ 48 hour	≥ 48 hour	-
<b>Chemical tested, Sterile Strip</b>			
EO sterilization temperatures 55-60 C° Time sterile by EO 7 hours	Brown	Brown	Pass
<b>Biological Test</b>			
EZ Test Biological Indicator	Brown	Brown	Pass
24 hour results for EO	Red	Red	Pass

To protect the privacy of customers or suppliers of Salamander Jewelry.  
We erased their address from this document.