

To protect the privacy of
customers or suppliers of
Salamander Jewelry.
We erased their address from
this document.

Document No. : TR-EO-004181016

Date : 18/10/2016

Page : 2 / 4

PRODUCT : ORDER NO. 2162744

SAMPLE SLIP PARAMETER

EXHAUST:04H DOWN TOTAL :11H ADMIN
PRETT :00H 33MIN STERI :07H DOWN
ACTUAL TIME

06:55:55 STEILIZATION SUCCESS

END	06:55:44	058,80	-005kPa
	06:55:44	058,80	-005kPa
	02:59:36	058,80	-015kPa
	02:49:14	058,80	000kPa
EXHAUST	02:48:44	058,80	000kPa
	02:39:14	058,70	000kPa
	02:29:14	058,90	000kPa
	02:19:14	058,90	-001kPa
	02:09:14	058,80	000kPa
	01:59:14	058,90	000kPa
	01:49:14	059,10	000kPa
	01:39:14	059,00	000kPa
	01:29:14	059,00	000kPa
	01:19:14	058,80	000kPa
	01:09:14	059,00	000kPa
	00:59:14	059,00	000kPa
	00:49:14	059,00	002kPa
	00:39:14	059,00	001kPa
	00:29:14	059,10	-001kPa
	00:19:14	059,10	000kPa
	00:09:14	059,10	000kPa
	23:59:14	059,00	000kPa
	23:49:14	059,00	-001kPa
	23:39:14	059,00	000kPa
	23:29:14	059,00	000kPa
	23:19:14	059,00	000kPa
	23:09:14	059,00	000kPa

	22:59:14	059,00	002kPa
	22:49:14	059,00	001kPa
	22:39:14	059,10	000kPa
	22:29:14	059,00	000kPa
	22:19:14	059,10	000kPa
	22:09:14	059,20	000kPa
	21:59:14	059,10	000kPa
	21:49:14	059,10	064kPa
	21:39:14	059,20	068kPa
	21:29:14	059,60	070kPa
	21:19:13	059,00	071kPa
	21:09:13	059,30	076kPa
	20:59:13	058,90	086kPa
	20:49:13	058,90	096kPa
	20:39:13	058,60	109kPa
	20:29:13	058,90	134kPa
	20:19:13	058,60	203kPa
	20:09:13	058,60	311kPa
	19:59:13	058,50	311kPa
	19:49:13	058,40	311kPa
STERI	19:48:56	058,50	311kPa
ADD EO	19:45:17	058,30	-088kPa
	19:39:13	058,00	-016kPa
HUMID1	19:35:15	057,90	-083kPa
PRETT	19:29:13	055,90	000kPa
PRHEAT	19:15:53	054,40	-001kPa
STATE	TIME	TEMP	PRE

RUN COUNT : 0319

EO PRE :060kPa VENT TIME:04H 00M

STERI TEMP: 056,00 STERI TIME:07H 00M

PRO TIP 19:15:53 054,40 -001kPa

START TIME: 2016-10-14 19:15:45