

To protect the privacy of customers or suppliers of Salamander Jewelry.
We erased their address from this document.

CERTIFICATE OF STERILIZATION

PRODUCT : **Order No. 1996428**

CERT. NO : CM 00135/12

CUSTOMER : Salamander Jewelry Co..Ltd

DATE : 18 /10 /2012

We here by certify that following products were sterilized by E.O.at
CITY MEDICAL SUPPLY. The result of sterilization by biological indicator are passed

CODE	Iteme	Color	LOT.NO	TOTAL QUANTITY	STERILIZED DATE	SPECIFICATION		RESULT		
						CHEMICAL TESTED	BIOLOGICAL TESTED	CHEMICAL TESTED	BIOLOGICAL TESTED	
Z-BC-1.2-10.0-4	STERILE SURGICAL STEEL BCR	CR	.000094	506	15 Oct 55	Positive	Negative	Positive	Negative	Pass
Z-BC-1.2-12.0-4	STERILE SURGICAL STEEL BCR			103						
Z-BC-1.6-10.0-4	STERILE SURGICAL STEEL BCR			507						
Z-BNDJ-1.6-10.0-5/8-CR	STERILE SURGICAL STEEL DOUBLE JEWELLED BANANA			138						
Z-LB-1.6-10.0-3	STERILE SURGICAL STEEL LABRET			224						
Z-MBN-1.2-10.0-3	STERILE SURGICAL STEEL MICRO BANANA			346						
Z-TBNJ-1.6-12.0-5/8-CR	STERILE TITANIUM JEWELLED BANANA			222						

To protect the privacy of customers or suppliers of Salamander Jewelry.
We erased their address from this document.