

To protect the privacy of  
customers or suppliers of  
Salamander Jewelry.  
We erased their address from  
this document.

Document No. : TR-EO-003160117

Date : 16/1/2017

Page : 2 / 4

PRODUCT : ORDER NO. 2171088

### SAMPLE SLIP PARAMETER

EXHAUST:04H 00MIN TOTAL :11H 57MIN  
PRETT :00H 52MIN STERI :07H 00MIN  
ACTUAL TIME

18:24:28 STEILIZATION SUCCESS

END 18:24:17 059,9C -005kPa  
18:24:17 059,9C -005kPa  
14:28:44 059,7C -015kPa  
EXHAUST 14:18:17 059,6C 000kPa  
14:16:47 059,6C -001kPa  
14:06:47 059,5C 000kPa  
13:56:47 059,4C 000kPa  
13:46:47 059,4C -001kPa  
13:36:47 059,5C 000kPa  
13:26:47 059,5C 000kPa  
13:16:47 059,3C 000kPa  
13:06:47 059,4C 000kPa  
12:56:47 059,3C 000kPa  
12:46:47 059,1C 000kPa  
12:36:47 059,2C 000kPa  
12:26:47 059,1C 001kPa  
12:16:47 059,1C 001kPa  
12:06:47 059,0C 000kPa  
11:56:47 058,8C -001kPa  
11:46:47 058,8C -001kPa  
11:36:47 058,7C -001kPa  
11:26:47 058,5C 002kPa  
11:16:47 058,6C 001kPa  
11:06:47 058,6C 000kPa  
10:56:47 058,4C 001kPa  
10:46:47 058,2C 000kPa  
10:36:47 058,2C 001kPa  
10:26:47 058,2C 001kPa

10:16:47 058,2C 001kPa  
10:06:47 058,2C 001kPa  
09:56:47 058,1C 001kPa  
09:46:47 057,8C 001kPa  
09:36:47 057,8C 001kPa  
09:26:47 057,5C 002kPa  
09:16:47 057,5C 001kPa  
09:06:47 057,5C 001kPa  
08:56:47 057,4C 000kPa  
08:46:47 057,3C 052kPa  
08:36:47 057,1C 056kPa  
08:26:47 057,2C 058kPa  
08:16:47 057,2C 065kPa  
08:06:47 057,0C 078kPa  
07:56:47 056,9C 101kPa  
07:46:47 057,0C 153kPa  
07:36:47 057,0C 300kPa  
07:26:47 057,0C 311kPa  
STERI 07:18:56 057,3C 252kPa  
07:16:47 057,3C 006kPa  
ADD EO 07:12:47 057,2C -087kPa  
07:06:47 057,2C -013kPa  
HUMIDI 07:02:45 057,1C -088kPa  
PRETT 06:56:47 057,2C 001kPa  
PREHEAT 06:27:27 057,1C -001kPa  
STATE TIME TEMP PRE

RUN COUNT : 1045  
EO PRE :060kPa VENT TIME:04H 00M  
STERI TEMP: 056,0C STERI TIME:07H 00M  
PRO TYP 06:27:27 057,1C -001kPa  
START TIME: 2017-01-13 06:27:19