

To protect the privacy of  
customers or suppliers of  
Salamander Jewelry.  
We erased their address from  
this document.

Document No. : TR-EO-006160217

Date : 16/2/2017

Page : 2 / 4

PRODUCT : ORDER NO. 2174693

### SAMPLE SLIP PARAMETER

EXHAUST:04H DOWN	TOTAL :12H DOWN	PRETT :01H 01MIN	STERI :07H 00MIN	ACTUAL TIME
06:00:24 STERILIZATION SUCCESS				
END	06:00:13	056,20	-004Pa	
	06:00:13	056,20	-004Pa	
	02:04:19	056,20	-013Pa	
	01:57:43	056,30	001kPa	
EXHAUST	01:57:13	056,30	001kPa	
	01:47:43	056,30	001kPa	
	01:37:43	056,30	001kPa	
	01:27:43	056,30	001kPa	
	01:17:43	056,40	001kPa	
	01:07:43	056,30	001kPa	
	00:57:43	056,30	001kPa	
	00:47:43	056,30	001kPa	
	00:37:43	056,50	001kPa	
	00:27:43	056,50	001kPa	
	00:17:43	056,60	001kPa	
	00:07:43	056,60	001kPa	
	23:57:43	056,60	001kPa	
	23:47:43	056,60	000kPa	
	23:37:43	056,60	001kPa	
	23:27:43	056,70	001kPa	
	23:17:43	056,60	001kPa	
	23:07:43	056,70	003kPa	
	22:57:43	056,70	001kPa	
	22:47:43	056,90	001kPa	
	22:37:43	056,90	003kPa	
	22:27:43	057,00	001kPa	
	22:17:43	056,90	002kPa	
	22:07:43	056,90	052kPa	
	21:57:43	056,90	054kPa	

	21:47:43	056,90	058kPa
	21:37:43	057,10	059kPa
	21:27:43	057,10	061kPa
	21:17:43	057,30	063kPa
	21:07:43	057,40	070kPa
	20:57:43	057,20	077kPa
	20:47:43	057,40	082kPa
	20:37:43	057,30	091kPa
	20:27:43	057,60	099kPa
	20:17:43	057,60	115kPa
	20:07:43	057,60	143kPa
	19:57:43	057,50	192kPa
	19:47:43	057,60	250kPa
	19:37:43	057,50	311kPa
	19:27:43	057,50	311kPa
	19:17:43	057,50	311kPa
	19:07:43	057,60	311kPa
	18:57:43	058,00	311kPa
STERI	18:57:20	058,00	311kPa
ACC EO	18:53:30	057,90	-008kPa
	18:47:43	057,70	-013kPa
HUMIDI	18:43:36	057,70	-088kPa
PRETT	18:37:43	057,50	002kPa
PREAT	17:56:11	052,30	002kPa
STATE	TIME	TEMP	PRE
RUN COUNT : 10X			
EO PRE	:050kPa	VENT TIME:04H 00M	
STERI TEMP:	056,00	STERI TIME:07H 00M	
PRO TIP	17:56:11	052,30	002kPa
START TIME: 2017-02-14 17:56:03			